

From: Cynthia Caporale/ESC/R3/USEPA/US
Sent: 3/22/2012 3:13:51 PM

To: [Ex. 4 - CBI]
CC: Kelley Chase/R3/USEPA/US@EPA
Subject: Re: Verification/Completeness Check for Dimock (R3 WO 1202005 PART 1 Posted Mar 15)

[Ex. 4 - CBI]

I'm ready to send the response on this one but wanted to check on the Cu blank issue. Are you raising the QL for the FB then putting a J on anything above 10X lik what was done for TDS?

Just putting a J on the result does not indicate enough about the blank contamination.

Cindy

Cynthia Caporale, Chief
OASQA Laboratory Branch
U.S. EPA Region III
Environmental Science Center
Fort Meade, MD
(410) 305-2732
Fax: (410) 305-3095

From: [Ex. 4 - CBI]
To: Cynthia Caporale/ESC/R3/USEPA/US@EPA, Kelley Chase/R3/USEPA/US@EPA
Cc: John Gilbert/CI/USEPA/US@EPA, Gary Newhart/CI/USEPA/US@EPA, [Ex. 4 - CBI]

Date: 03/21/2012 11:29 AM
Subject: Verification/Completeness Check for Dimock (R3 WO 1202005 PART 1 Posted Mar 15)

.....is attached for your review and consideration.

[Ex. 4 - CBI]

Lockheed Martin

Scientific, Engineering, Response and Analytical Services (SERAS)

[Ex. 4 - CBI]

[Ex. 4 - CBI]

[attachment "SERAS-172-DSR-032112_32.docx" deleted by Cynthia Caporale/ESC/R3/USEPA/US]